# **Purge Assemblies**

## **FPV-1 Series**

### **Technical Data**

- O Maximum working pressure: 4500 psig
- Material of the main components:
  Seat: PCTFE (diaphragm valve)
  Diaphragm: cobalt alloy (diaphragm valve)
- $\bigcirc$  Temperature: -10 °F ~ 150 °F (-23 °C ~ 65 °C)
- Leak rates (helium): Internal: ≤1x10<sup>.9</sup> std cm<sup>3</sup>/s External: ≤1x10<sup>.9</sup> std cm<sup>3</sup>/s
- ◎ Minimum orifice:  $\Phi$ 0.13" (3.2 mm)

#### **Product Types**

- Straight Purge Assemblies
  - Consisting of a diaphragm valve and a check valve (see Fig. 1-1).

Connecting the auxiliary inlet port (see Fig. 1-2) of the regulator or in between the regulator and the cylinder (see Fig. 1-3) to allow the corrosive or toxic gas to be vented through to a safe location.



O Tee Purge Assemblies

FITOK

Consisting of a diaphragm valve, check valve, tee, and cylinder connections (see Fig. 2-1). Connecting the cylinder with the regulator. Before installing a new cylinder, open the diaphragm valve, and the remaining gas is vented safely; or after a new cylinder is installed, close the regulator and open the diaphragm valve, enabling the process gas inside the cylinder to purge the atmospheric contamination between the cylinder and the regulator (see Fig. 2-2).



#### ◎ Cross Purge Assemblies

Consisting of a tee purge assembly and two additional diaphragm valves (see Fig. 3-1).

Except for process gas, purging is also made possible with inert gas from outside (see Fig. 3-2). The steps are as follows: Before installing a new cylinder, close the diaphragm valve beside the regulator, and open the shutoff diaphragm valve on the vent line to release the residual pressure.

After installing a new cylinder, open the diaphragm valve on the bottom, allowing the inert gas from outside to purge the atmospheric contaminations between the cylinder and the diaphragm valve.



#### **Part Number Description**

	FPV - 1C		6L	-	DIN1 -	_	00 -	- 00	- 00
Product Type		Body Material			Inlet Port		Vent Port	Outlet Port	Purge Inlet
s	Straight Purge Assemblies	6L	. 316L SS	00	1/4" Female NPT	00	1/4" Female NPT	Same as Vent Port	Same as Vent Port
т	Tee Purge	SS	316 SS	01	1/4" Male NPT	01	1/4" Male NPT		
	Cross Purge			C	CGA Number (USA)		1/4" Tube		
C	Assemblies			DIN_	DIN Number (Germany)	10	Fitting		
						20	6 mm Tube		
				Refer to cylinde	to page B-30 for er connections	21	8 mm Tube		
				based Cylind compli standa upon r contac	based on specific gas type. Cylinder connections compliant to other standards are available upon request. Please contact FITOK Group for		ner connections available upon juest		